

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)




AT-1807

Laboratory Report

for

West Slope Water District
31PO Box 25140
Portland, OR 97298
Attention: Michael Grimm



UTAH ELCP CA00006

Report:838770
Project:UCMR4
Group:AM3-EP-OR4100660-West Slope Water District

ZIA8: Vanessa Berry
Project Manager

- * Accredited in accordance with TNI 2016 and ISO/IEC 17025:2017.
- * Laboratory certifies that the test results meet all **TNI 2016 and ISO/IEC 17025:2017** requirements unless noted under the individual analysis.
- * Following the cover page are State Certification List, ISO/IEC 17025:2017 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms, as applicable.
- * Test results relate only to the sample(s) tested.
- * Test results apply to the sample(s) as received, unless EEA-M collected and analyzed the sample(s) as noted in the COC and final report.
- * This report shall not be reproduced except in full, without the written approval of the laboratory.
- * This report includes ISO/IEC 17025:2017 and non-ISO/IEC 17025:2017 accredited methods.

STATE CERTIFICATION LIST

State	Certification Number	State	Certification Number
Alabama	41060	Montana	Cert 0035
Arizona	AZ0778	Nebraska	Certified
Arkansas	Certified	Nevada	CA000062018
California	2813	New Hampshire *	2959
Colorado	Certified	New Jersey *	CA 008
Connecticut	PH-0107	New Mexico	Certified
Delaware	CA 006	New York *	11320
Florida *	E871024	North Carolina	06701
Georgia	947	North Dakota	R-009
Guam	18-005R	Oregon *	CA200003-005
Hawaii	Certified	Pennsylvania *	68-565
Idaho	Certified	Puerto Rico	Certified
Illinois *	200033	Rhode Island	LAO00326
Indiana	C-CA-01	South Carolina	87016
Iowa - Asbestos	413	South Dakota	Certified
Kansas *	E-10268	Tennessee	TN02839
Kentucky	90107	Texas *	T104704230-18-15
Louisiana *	LA180000	Utah (Primary AB) *	CA00006
Maine	CA0006	Vermont	VT0114
Maryland	224	Virginia *	460260
Commonwealth of Northern Marianas Is.	MP0004	Washington	C838
Massachusetts	M-CA006	EPA Region 5	Certified
Michigan	9906	Los Angeles County Sanitation Districts	10264
Mississippi	Certified		

* NELAP/TNI Recognized Accreditation Bodies

ISO/IEC 17025 Accredited Method List

The tests listed below are accredited and meet the requirements of ISO/IEC 17025 as verified by the ANSI-ASQ National Accreditation Board/ANAB. Refer to Certificate and scope of accreditation (AT 1807) found at: <https://www.eurofinsus.com/Eaton>

SPECIFIC TESTS	METHOD OR TECHNIQUE USED	Environmental (Drinking Water)	Environmental (Waste Water)	Water as a Component of Food and Bev/Bev/ Bottled Water
1,2,3-TCP (5 PPT & 0.5 PPT)	CA SRL 524M-TCP	x		x
1,4-Dioxane	EPA 522	x		x
2,3,7,8-TCDD	Modified EPA 1613B	x		x
Acrylamide	In House Method (2440)	x		x
Algal Toxins/Microcystin	In House Method (3570)			
Alkalinity	SM 2320B	x	x	x
Ammonia	EPA 350.1		x	x
Ammonia	SM 4500-NH3 H		x	x
Anions and DBPs by IC	EPA 300.0	x	x	x
Anions and DBPs by IC	EPA 300.1	x		x
Asbestos	EPA 100.2	x	x	
BOD / CBOD	SM 5210B		x	x
Bromate	In House Method (2447)	x		x
Carbamates	EPA 531.2	x		x
Carbonate as CO3	SM 2330B	x	x	x
Carbonyls	EPA 556	x		x
COD	EPA 410.4 / SM 5220D		x	
Chloramines	SM 4500-CL G	x	x	x
Chlorinated Acids	EPA 515.4	x		x
Chlorinated Acids	EPA 555	x		x
Chlorine Dioxide	SM 4500-CLO2 D Palin Test	x		x
Chlorine -Total/Free/ Combined Residual	SM 4500-CI G	x	x	x
Conductivity	EPA 120.1		x	
Conductivity	SM 2510B	x	x	x
Corrosivity (Langelier Index)	SM 2330B	x		x
Cyanide, Amenable	SM 4500-CN G	x	x	
Cyanide, Free	SM 4500CN F	x	x	x
Cyanide, Total	EPA 335.4	x	x	x
Cyanogen Chloride (screen)	In House Method (2470)	x		x
Diquat and Paraquat	EPA 549.2	x		x
DBP/HAA	SM 6251B	x		x
Dissolved Oxygen	SM 4500-O G		x	x
DOC	SM 5310C	x		x
E. Coli	(MTF/EC+MUG)	x		x
E. Coli	CFR 141.21(f)(6)(i)	x		x
E. Coli	SM 9223		x	
E. Coli (Enumeration)	SM 9221B.1/ SM 9221F	x		x
E. Coli (Enumeration)	SM 9223B	x		x
EDB/DCBP	EPA 504.1	x		
EDB/DCBP and DBP	EPA 551.1	x		x
EDTA and NTA	In House Method (2454)	x		x
Endothall	EPA 548.1	x		x
Endothall	In-house Method (2445)	x		x
Enterococci	SM 9230B	x	x	
Fecal Coliform	SM 9221 E (MTF/EC)	x		
Fecal Coliform	SM 9221C, E (MTF/EC)		x	
Fecal Coliform (Enumeration)	SM 9221E (MTF/EC)	x		x
Fecal Coliform with Chlorine Present	SM 9221E		x	
Fecal Streptococci	SM 9230B	x	x	
Fluoride	SM 4500-F C	x	x	x
Glyphosate	EPA 547	x		x
Glyphosate + AMPA	In House Method (3618)	x		x
Gross Alpha/Beta	EPA 900.0	x	x	x
Gross Alpha Coprecipitation	SM 7110 C	x	x	x
Hardness	SM 2340B	x	x	x
Heterotrophic Bacteria	In House Method (2439)	x		x
Heterotrophic Bacteria	SM 9215 B	x		x
Hexavalent Chromium	EPA 218.6	x	x	x

SPECIFIC TESTS	METHOD OR TECHNIQUE USED	Environmental (Drinking Water)	Environmental (Waste Water)	Water as a Component of Food and Bev/Bev/ Bottled Water
Hexavalent Chromium	EPA 218.7	x		x
Hexavalent Chromium	SM 3500-Cr B		x	
Hormones	EPA 539	x		x
Hydroxide as OH Calc.	SM 2330B	x		x
Kjeldahl Nitrogen	EPA 351.2		x	
Legionella	Legiolert	x		x
Mercury	EPA 245.1	x	x	x
Metals	EPA 200.7 / 200.8	x	x	x
Microcystin LR	ELISA (2360)	x		x
Microcystin, Total	EPA 546	x		x
NDMA	EPA 521 In house method (2425)	x		x
Nitrate/Nitrite Nitrogen	EPA 353.2	x	x	x
OCL, Pesticides/PCB	EPA 505	x		x
Ortho Phosphate	EPA 365.1	x	x	x
Ortho Phosphorous	SM 4500P E	x		x
Oxyhalides Disinfection Byproducts	EPA 317.0	x		x
Perchlorate	EPA 331.0	x		x
Perchlorate (low and high)	EPA 314.0	x		x
Perfluorinated Alkyl Acids	EPA 537	x		x
Perfluorinated Pollutant	In house Method (2434)	x		x
pH	EPA 150.1	x		
pH	SM 4500-H+B	x	x	x
Phenylurea Pesticides/ Herbicides	In House Method, based on EPA 532 (2448)	x		x
Pseudomonas	IDEXX Pseudalart (2461)	x		x
Radium-226	GA Institute of Tech	x		x
Radium-228	GA Institute of Tech	x		x
Radon-222	SM 7500RN	x		x
Residue, Filterable	SM 2540C	x	x	x
Residue, Non-filterable	SM 2540D		x	
Residue, Total	SM 2540B		x	x
Residue, Volatile	EPA 160.4		x	
Semi-VOC	EPA 525.2	x		x
Silica	SM 4500-Si D	x	x	
Silica	SM 4500-SiO2 C	x	x	
Sulfide	SM 4500-S ⁻ D		x	
Sulfite	SM 4500-SO ³ B	x	x	x
Surfactants	SM 5540C	x	x	x
Taste and Odor Analytes	SM 6040E	x		x
Total Coliform (P/A)	SM 9221 A, B	x		x
Total Coliform (Enumeration)	SM 9221 A, B, C	x		x
Total Coliform / E. coli	Colisure SM 9223	x		x
Total Coliform	SM 9221B		x	
Total Coliform with Chlorine Present	SM 9221B		x	
Total Coliform / E.coli (P/A and Enumeration)	SM 9223	x		x
TOC	SM 5310C	x	x	x
TOX	SM 5320B		x	
Total Phenols	EPA 420.1		x	
Total Phenols	EPA 420.4	x	x	x
Total Phosphorous	SM 4500 P E		x	
Triazine Pesticides & Degradates	In House (3617)	x		x
Turbidity	EPA 180.1	x	x	x
Turbidity	SM 2130B	x	x	
Uranium by ICP/MS	EPA 200.8	x		x
UV 254	SM 5910B	x		
VOC	EPA 524.2	x		x
VOC	In House Method (2411)	x		x
Yeast and Mold	SM 9610	x		x
Field Sampling	N/A			

Acknowledgement of Samples Received

Addr: **West Slope Water District**
31PO Box 25140
Portland, OR 97298

Attn: Michael Grimm
Phone: 503-292-2777

Client ID: WESTSLOPE-OR
Folder #: 838770
Project: UCMR4
Sample Group: AM3-EP-OR4100660-West Slope
Water District
Project Manager: Vanessa Berry
Phone: 503-310-3905

The following samples were received from you on **November 14, 2019 at 1207**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, LLC.

Sample #	Sample ID	Sample Date
201911140404	11940-EP001-Entry Point to Dist. System	11/13/2019 1245
	Sample Type: EP Sample Event: SEC7 Facility ID: 11940 Sample Point ID: EP001 PWSID: OR4100660	
	@UCMR4 544 HOLD	@UCMR4 545
		@UCMR4 546

Test Description

- @UCMR4 544 HOLD -- @UCMR4 544 HOLD
- @UCMR4 545 -- UCMR4 545
- @UCMR4 546 -- UCMR4 546



Eaton Analytical

750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016-3629

Phone: 626 386 1100/ 800 566 LABS (800 566 5227)
Fax: 626 386 1101

Website: <http://www.eurofinsus.com/Eaton>

UCMR4 CHAIN OF CUSTODY RECORD

EUROFINS EATON ANALYTICAL USE ONLY:

Folder No:

LOGIN COMMENTS:

SAMPLES CHECKED AGAINST COC BY:

SAMPLES LOGGED IN BY:

SAMPLE TEMP RECEIVED Criteria: (recorded in internal COC)

SAMPLES REC'D DAY OF COLLECTION?

(check for yes)

If the sample(s) received:

- same day receipt as sample collection with evidence of cooling, sample temperature ≥ 10°C is acceptable
- within the first 48 hours of collection time; sample temperature must be ≤ 10°C (except 200.8) and not frozen (except 546), and
- after 48 hours of collection time; sample temperature must be ≤ 6°C (except 200.8) and not frozen (except 546), and valid if refrigerated between collection and shipment documented below as "yes."

PWSID: OR.00660
 Example: (CA1234567)

For PWS and Intermediate Lab Use ONLY:
 Were samples cooled between sample collection and shipment at 10°C, or less, for the first 48 hours and 6°C, or less, thereafter? If yes, please "✓" the box next to yes below.
 No documentation of cooling of samples between collection and shipment for samples received after 48 hours of sample collection will be rejected.
 Intermediate Lab: YES NO

TO BE COMPLETED BY SAMPLER:

(check for yes)

COMPANY/AGENCY NAME: West Slope Water Dist. PROJECT CODE: UCMR4

EEA CLIENT CODE: WEST Slope Water Dist. SAMPLE GROUP: UCMR4

NOTE: we MUST have PWSID#, Facility ID, Sample Point ID, and Sample event # to be able to upload data to EPA Database (check for yes). OR

TAT requested: rush by adv notice only	FACILITY ID (per EPA Requirement) - 5 characters Max	SAMPLE POINT ID (per EPA Requirement) - 20 characters max	SAMPLE EVENT #	STD	1 wk	3 day	2 day	1 day	SAMPLE COMMENTS
11-13-19 12:45	Entry Point	UCMR4 544							
↓	↓	UCMR4 544							
↓	↓	UCMR4 545							
		UCMR4 546							
		UCMR4 200.8							
		UCMR4 525.3							
		UCMR4 530							
		UCMR4 541							
		UCMR4 552.3							
		TOC							
		Bromide							

(1) Sample Event Code: Cyanotoxins -- SEC1 SEC2 SEC3 SEC4 SEC5 SEC6 SEC7 SEC8
 HAAs -- SEH1 SEH2 SEH3 SEH4
 Metals, Pesticides, Alcohols, SVOCs -- SEA1 SEA2 SEA3 SEA4

SIGNATURE: Bart Johnson PRINT NAME: Bart Johnson COMPANY: West Slope Water Dist. DATE: 11-13-19 TIME: 12:45

RELINQUISHED BY: Bart Johnson RECEIVED BY: Jason O'Conner DATE: 11/13/19 TIME: 1330

RELINQUISHED BY: Bart Johnson RECEIVED BY: Bart Johnson DATE: 11-13-19 TIME: 1350

QA EO 0103.1 (QA FO-FRM17289) Version 2 (12/28/2017)

Relinquished by Vanessa Berry RECEIVED BY: Vanessa Berry

11-13-19 1330



Eaton Analytical

UCMR4 INTERNAL CHAIN OF CUSTODY RECORD

EEA Folder Number: 8357A1

SAMPLES RECEIVED WITHIN 48 HOURS OF COLLECTION TIME?

TYPE OF ICE: Real Synthetic No Ice

CONDITION OF ICE: Frozen Partially Frozen Thawed N/A

CONDITION OF SAMPLE: Frozen Partially Frozen Not Frozen

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: _____

Compliance Acceptance Criteria:

If sample(s) received:

105638692053

- 1) on the same day as the collection day; sample temperature may be $\geq 10^{\circ}\text{C}$ with evidence of cooling
- 2) within the first 48 hours of collection time; sample temperature must be $\leq 10^{\circ}\text{C}$ (except 200.8) and not frozen (except 546), and
- 3) after 48 hours of collection time; sample temperature must be $\leq 6^{\circ}\text{C}$ (except 200.8) and not frozen (except 546), and not rejected if refrigerated between collection and shipment documented on UCMR4 COC as "yes."

Note: A minimum of 1 bottle for every analytical method must be checked for temperature. If the bottle that is checked does not meet the temperature criterion, then the sample bottle is rejected. The temperature of the other samples collected for that method is checked to determine if a valid sample was received.

Facility ID & Unique Field Sample ID

11940-EP001

IR Gun ID = 6160

Method	Container ID	Observation (°C)	Correction Factor (°C)	Final (°C)
UCMR4 2008	1		+	=
UCMR4 525.3	1		+	=
	2		+	=
	3		+	=
UCMR4 530	1		+	=
	2		+	=
	3		+	=
UCMR4 541	1		+	=
	2		+	=
	3		+	=
UCMR4 552.3	1		+	=
TOC (5310C)	1		+	=
Bromide (300.0)	1		+	=

Method	Container ID	Observation (°C)	Correction Factor (°C)	Final (°C)
UCMR4 544	1	3.0	-0.3	= 2.7
	2		+	=
	3		+	=
UCMR4 545	1	3.7	-0.3	= 3.4
UCMR4 546	1	4.6	-0.3	= 4.3

Note: If samples are out of temperature range, let the ASM's know. ASM's will determine whether to proceed with analysis or not.

SIGNATURE	PRINT NAME	COMPANY/TITLE	DATE	TIME
<u>Chuck Brooks</u>	<u>Chuck Brooks</u>	<u>Eurofins Eaton Analytical</u>	<u>11.14.19</u>	<u>1207</u>

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Comments

Report: 838770
Project: UCMR4
Group: AM3-EP-OR4100660-West Slope
Water District

West Slope Water District
Michael Grimm
31PO Box 25140
Portland, OR 97298

Tel: (626) 386-1100
Fax: (626) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Hits

Report: 838770
Project: UCMR4
Group: AM3-EP-OR4100660-West Slope
Water District

West Slope Water District

Michael Grimm
31PO Box 25140
Portland, OR 97298

Samples Received on:
11/14/2019 1207

Analyzed	Analyte	Sample ID	Result	Federal MCL	Units	MRL
----------	---------	-----------	--------	-------------	-------	-----

Tel: (626) 386-1100
 Fax: (866) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 838770
Project: UCMR4
Group: AM3-EP-OR4100660-West Slope Water District

West Slope Water District
 Michael Grimm
 31PO Box 25140
 Portland, OR 97298

Samples Received on:
 11/14/2019 1207

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
11940-EP001-Entry Point to Dist. System (201911140404)						Sampled on 11/13/2019 1245			
Sample Type: EP									
Sample Event: SEC7									
Facility ID: 11940									
Sample Point ID: EP001									
PWSID: OR4100660									
EPA 546 - UCMR4 546									
11/14/19	11/15/19 13:58	1209805	1210130	(EPA 546)	Total Microcystins	ND	ug/L	0.30	1
11/14/19	11/15/19 13:58	1209805	1210130	(EPA 546)	%CV	0.200	%	20.0	1
EPA 545 - UCMR4 545									
11/19/19	11/21/19 2:20	1207262	1211308	(EPA 545)	Anatoxin-a	ND	ug/L	0.030	1
11/19/19	11/20/19 1:50	1207262	1210848	(EPA 545)	Cylindrospermopsin	ND	ug/L	0.090	1
11/19/19	11/21/19 2:20	1207262	1211308	(EPA 545)	L-Phenylalanine-d5	88	%		1
11/19/19	11/20/19 1:50	1207262	1210848	(EPA 545)	Uracil-d4	93	%		1
EPA 544 - @UCMR4 544 HOLD									
	11/15/19 13:58		1210152	(EPA 544)	Prep Information	NA	ug/L		1

Rounding on totals after summation.
 (c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Report: 838770
Project: UCMR4
Group: AM3-EP-OR4100660-West Slope
Water District

West Slope Water District

UCMR4 546

Prep Batch: 1209805 Analytical Batch: 1210130

201911140404 11940-EP001-Entry Point to Dist. System

Analysis Date: 11/15/2019

Analyzed by: R77L

@UCMR4 544 HOLD

Analytical Batch: 1210152

201911140404 11940-EP001-Entry Point to Dist. System

Analysis Date: 11/15/2019

Analyzed by: R77L

UCMR4 545

Prep Batch: 1207262 Analytical Batch: 1210848

201911140404 11940-EP001-Entry Point to Dist. System

Analysis Date: 11/20/2019

Analyzed by: SZZ

UCMR4 545

Prep Batch: 1207262 Analytical Batch: 1211308

201911140404 11940-EP001-Entry Point to Dist. System

Analysis Date: 11/21/2019

Analyzed by: SZZ

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 838770
 Project: UCMR4
 Group: AM3-EP-OR4100660-West Slope
 Water District

West Slope Water District

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPDLimit (%)	RPD%
---------	---------	--------	--------	-----------	-------	-----------	------------	--------------	------

UCMR4 546 by EPA 546

Analytical Batch: 1210130

Analysis Date: 11/15/2019

LCS1	%CV			2.20	%				
LCS2	%CV			ND	%				
MBLK	%CV			<15	%				
MBLK	%CV			<15	%				
MRL_CHK	%CV			ND	%				
MS2_201911080196	%CV	2.90		ND		15			
MSD2_201911080196	%CV	2.90		ND		15			
LCS1	Total Microcystins		0.5	0.427	ug/L	85	(60-140)		
LCS2	Total Microcystins		0.5	0.475	ug/L	95	(60-140)		
MBLK	Total Microcystins			<0.15	ug/L				
MBLK	Total Microcystins			<0.15	ug/L				
MRL_CHK	Total Microcystins		0.3	0.244	ug/L	81	(50-150)		
MS2_201911080196	Total Microcystins	ND	0.5	0.184	ug/L	<u>30</u>	(60-140)		
MSD2_201911080196	Total Microcystins	ND	0.5	0.257	ug/L	<u>44</u>	(60-140)	40	33

UCMR4 545 by EPA 545

Analytical Batch: 1210848

Analysis Date: 11/19/2019

CCCH	Cylindrospermopsin		3	3.17	ug/L	106	(70-130)		
CCCL	Cylindrospermopsin		0.09	0.0840	ug/L	93	(50-150)		
CCCM	Cylindrospermopsin		0.9	0.934	ug/L	104	(70-130)		
MBLK	Cylindrospermopsin			<0.030	ug/L				
MRL_CHK	Cylindrospermopsin		0.09	0.0840	ug/L	93	(50-150)		
MS_201911050527	Cylindrospermopsin	ND	0.09	0.0660	ug/L	73	(50-150)		
MSD_201911050527	Cylindrospermopsin	ND	0.09	0.0770	ug/L	86	(50-150)	50	15
CCCH	Uracil-d4 (S)		100	96.2	%	96	(50-150)		
CCCL	Uracil-d4 (S)		100	88.8	%	89	(50-150)		
CCCM	Uracil-d4 (S)		100	95.0	%	95	(50-150)		
MBLK	Uracil-d4 (S)			87.0	%	87	(50-150)		
MRL_CHK	Uracil-d4 (S)		100	88.8	%	89	(50-150)		
MS_201911050527	Uracil-d4 (S)		100	93.2	%	93	(50-150)		
MSD_201911050527	Uracil-d4 (S)		100	88.6	%	89	(50-150)		

UCMR4 545 by EPA 545

Analytical Batch: 1211308

Analysis Date: 11/20/2019

CCCH	Anatoxin-a		1	1.01	ug/L	101	(70-130)		
CCCL	Anatoxin-a		0.03	0.0200	ug/L	67	(50-150)		

Spike recovery is already corrected for native results.
 Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.
 Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.
 RPD not calculated for LCS2 when different a concentration than LCS1 is used.
 RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).
 (S) - Indicates surrogate compound.
 (I) - Indicates internal standard compound.

Tel: (626) 386-1100
 Fax: (626) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 838770
Project: UCMR4
Group: AM3-EP-OR4100660-West Slope
 Water District

West Slope Water District

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPDLimit (%)	RPD%
CCCL	Anatoxin-a		0.03	0.0180	ug/L	60	(50-150)		
CCCM	Anatoxin-a		0.3	0.283	ug/L	94	(70-130)		
CCCM	Anatoxin-a		0.3	0.270	ug/L	90	(70-130)		
MBLK	Anatoxin-a			<0.010	ug/L				
MRL_CHK	Anatoxin-a		0.03	0.0200	ug/L	67	(50-150)		
MS_201911050527	Anatoxin-a	ND	0.03	0.0330	ug/L	110	(50-150)		
MSD_201911050527	Anatoxin-a	ND	0.03	0.0340	ug/L	113	(50-150)	50	3.0
CCCH	Cylindrospermopsin		3	3.77	ug/L	126	(70-130)		
CCCH	L-Phenylalanine-d5 (S)		100	95.0	%	95	(50-150)		
CCCL	L-Phenylalanine-d5 (S)		100	101	%	101	(50-150)		
CCCL	L-Phenylalanine-d5 (S)		100	146	%	146	(50-150)		
CCCM	L-Phenylalanine-d5 (S)		100	102	%	102	(50-150)		
CCCM	L-Phenylalanine-d5 (S)		100	79.3	%	79	(50-150)		
MBLK	L-Phenylalanine-d5 (S)			107	%	107	(50-150)		
MRL_CHK	L-Phenylalanine-d5 (S)		100	101	%	101	(50-150)		
MS_201911050527	L-Phenylalanine-d5 (S)		100	81.6	%	82	(50-150)		
MSD_201911050527	L-Phenylalanine-d5 (S)		100	95.3	%	95	(50-150)		
CCCH	Uracil-d4 (S)		100	88.6	%	89	(50-150)		

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.